

# DIGITAL INSULATION/CONTINUITY TESTER KEW 3551 / 3552 / 3552BT



#### **LED light & Display backlight**

Facilitate working at dimly illuminated location.

Illuminance sensor turns the LCD backlight and LED spot light ON/OFF.





#### **Diagnostic Insulation Tests**

Polarization Index

Insulation resistance value 10 min. after start Insulation resistance value 1 min. after start 4.0 or more 4.0 to 2.0

DAR

Dielectric Absorption Ratio									
DAR=-	Insulation resistance value 1 min. after start								
	nsulation resista	nce value 15 sec	. after start						
DAR	1.4 or more	1.25 to 1.0	1.0 or less						
Critoria	Reet	Good	Rad						

#### Memory/data transfer function (available on KEW3552/3552BT)

Internal memory up to 1000 measurements can be transferred to a PC by the optional adapter 8212-USB.

#### Bluetooth® communication (KEW 3552BT)

Transfer measured data directly to the compatible tablet or Smart phone devices.

Save time - easy to create reports.

Available on download site for free.



\*Please contact us with inquiries about the purchase of 3552BT.

KEW 3551 / 3552 / 3552BT Specification

### Provide maximum versatility with the following functions.

- Insulation resistance measurement (6 ranges)
- AC/DC voltage measurement
- Continuity test @200mA, Low-resistance



#### The right model to meet your needs

	Model	6 ranges insulation resistance	Memory	PC communi- cation	Bluetooth®	
	KEW 3551	•				
	KEW 3552	•	•	•		
	KEW 3552BT	•	•	•	•	

## Accessories



Insulation resistance							
Test voltage	50V	100V	125V	250V	500V	1000V	
Measuring range (Auto-ranging)	4.000/40.00/ 100.0MΩ	4.000/40.00/ 200.0MΩ	4.000/40.00/ 250.0MΩ	4.000/40.00/ 400.0/500.0MΩ	4.000/40.00/ 400.0/2000MΩ /20GΩ* <sup>1</sup>	4.000/40.00/ 400.0/4000MΩ /40GΩ* <sup>1</sup>	
Mid-scale value	2ΜΩ 5ΜΩ		10ΜΩ	100ΜΩ	200ΜΩ		
First effective measuring range	0.100 to 10.00MΩ	0.100 to 20.00M $\Omega$	0.100 to 25.00MΩ	0.100 to 50.0MΩ	0.100 to 500MΩ	0.100 to 1000MΩ	
Accuracy	±2%rdg±2dgt						
Second effective measuring range	0.050 to 0.099MΩ						
Second effective measuring range	10.01 to 100.0MΩ	20.01 to 200.0M $\Omega$	25.01 to 250.0MΩ	50.1 to 500MΩ	501 to 2000MΩ	1001 to 4000MΩ	
Accuracy	±5%rdg	±5%rdg					
Rated current	1.0 to 1.1mA						
nated current	@0.05MΩ	@0.1MΩ	@0.125MΩ	@0.25MΩ	@0.5MΩ	@1MΩ	
Output short circuit current	1.5mA max						
Ω/Continuity*3							
Resistance range	40.00/400.0/4000Ω						
Accuracy	±2.5%rdg±8dgt						
Open-circuit voltage 5V(4 to 6.9V)							
Measuring current 200mA or more							
Voltage							
Range AC 2.0 to 600V(45 to 65Hz)DC -2.0 to -600V +2.0 to +600V							
Accuracy	±1%rdg±4dgt						
General							
Applicable standards	IEC 61010 CAT III 600V / CAT IV 300V IEC 61010-2-034, IEC 61557-1,2,4, IEC 61326-1,2-2, IEC 60529(IP40)						
Communication interface	USB*1, Bluetooth® *2						
Dimension / Weight	97(L)x156(W)x46(D)mm / Approx. 490g(including battery)						
Power source	LR6/R6(AA)(1.5V) x 4						
Accessories	7260(Test lead with remote control switch), 7261A(Test lead with alligator clip), 8017A(Extension prod long), 9173(Carrying case), 9121(Shoulder strap), LR6(AA)x4, Instruction manual						
Optional accessories	7243A(L-shaped probe), 8016(Hook type prod), 8212-USB(USB adapter with "KEW Report(Software)")*1, 9186A(Carrying case), 9187(Cord case)						

- \*1 3552/3552BT only \*2 3552BT only, Bluetooth\* is a trademark or registered trademark of Bluetooth SIG, Inc. \*3 Low-resistance range is protected by a built-in fuse (0.5 A/ 1000 V, Dia. 6.3 x 32 mm)

#### **Optional Accessories**



Please read the "Safety Warnings" in the instruction manual supplied with the instrument thoroughly and completely Safety Warnings: for correct use. Failure to follow the safety rules can cause fire, trouble, electrical shock, etc. Therefore, make sure to operate the instrument on a correct power supply and voltage rating marked on each instrument.

## For inquiries or orders:



# KYORITSU ELECTRICAL **INSTRUMENTS** WORKS, LTD.

2-5-20, Nakane, Meguro-ku, Tokyo, 152-0031 Japan Phone:+81-3-3723-0131 Fax:+81-3-3723-0152

www.kew-ltd.co.jp

